



Schematic Diagram of sealing provision of the model.

Sealing point is provided on the two diagonal sides incorporate the stamping plate. A typical schematic diagram of sealing provision of the model is given above.

A dip switch is provided to disable the access the external calibration.

Further, in exercise of the power conferred by sub-rule (9) of rule 8 of the Legal Metrology (Approval of Models) Rules, 2011, the Central Government hereby declares that this certificate of approval of the said model shall also cover the weighing instruments of similar make, accuracy and performance of same series with maximum capacity up to 50kg with number of verification scale interval (n) in the range of 100 to 10000 for 'e' value of 100mg to 2g and with number of verification scale interval (n) in the range of 500 to 10000 for 'e' value of 5g or more and with 'e' value of  $1 \times 10^k$ ,  $2 \times 10^k$  or  $5 \times 10^k$ , where k is a positive or negative whole number or equal to zero manufactured by the same manufacturer in accordance with the same principle, design and with the same materials with which, the said approved model has been manufactured

[F.No.WM-21(232)/2014]

*B. N. Dixit*

( B. N. Dixit)

Director ( Legal Metrology) to Government of India

*Ashutosh Agarwal*  
 आशुतोष अग्रवाल / ASHUTOSH AGARWAL  
 सहा. निदेशक (मिथिक माप विज्ञान) / Assistant Director (Legal Metrology)  
 भारत सरकार / Government of India  
 उपभोक्ता मामले विभाग / Deptt. of Consumer Affairs  
 उपभोक्ता खाद्य एवं सार्वजनिक वितरण मंत्रालय  
 M/o Consumer Affairs, Food & P.D.  
 कृषि भवन, नई दिल्ली / Krishi Bhawan, New Delhi - 01