

**Notification**

S.O.--- Whereas the Central Government, after considering the report submitted to it by the prescribed authority, is satisfied that the model described in the said report (see the figure given below) is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions:

Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, the Central Government hereby issues and publishes the certificate of approval of the model of non-automatic weighing instrument of medium accuracy (accuracy class-III) of series "SSW" and with brand name "SIVA SCALES" (hereinafter referred to as the said model), manufactured by M/s Siva Scales, 29/A, Chandaka Industrial Area, Patia, Bhubaneswar-751024 (Odisha) and which is assigned the approval mark IND/09/14/ 446

The said model is a strain gauge type load cell based non-automatic weighing instrument with a maximum capacity of 60 tonne and minimum capacity of 200 kg. The verification scale interval (e) is 10 kg. It has subtractive tare. The Light Emitting Diode indicates the weighing result. The instrument operates on 230Volts, 50Hertz alternating current power supply. It is used as Electronic Weighbridge type weighing instrument.

Figure

*Ashutosh Agarwal*  
Ashutosh Agarwal / ASHUTOSH AGARWAL  
अशुतोष अग्रवाल / अशुतोष अग्रवाल  
सद. निदेशक (विशेष) एवं निदेशक / Additional Director (Legal Metrology)  
नारायण चरणचन्द्र / Government of India  
उपनिवेशता मंत्रालय / Ministry of Consumer Affairs  
उपनिवेशता खाद्य एवं सार्वजनिक वितरण मंत्रालय  
Min. Consumer Affairs, Food & P.D.  
एन.डी. बिल्डिंग, इंदिरा भवन, नई दिल्ली-110001  
Indira Bhawan, New Delhi-01

